



CONFORMITY CERTIFICATE OF EXPLOSION-PROOF

Certificate No.: CCRI 23.7628

Applicant Company: Beijing Dorland System Control Tech Co.,Ltd.
Address: Room1107,Tower B,Tsinghua Tongfang Technology Square,No.1 Wangzhuang Road,Beijing,P.R.China
Name of Product: Intrinsically safe infrared thermometer
Type of Product: EX-DT80
Marking: Ex ib II C T4 Gb/Ex ib III C T130°C Db
Technical Documents: Q/DLM 010-2023
Drawing No.: ZZ01
Note(s):
1.Rated voltage DC 9V, powered by a single 6F22 dry battery.
2.Operating environment temperature: $0^{\circ}\text{C} \leq T_a \leq +50^{\circ}\text{C}$.
3.IP code:IP54.
4.Correlation test report number:202355908.

By verifying the drawings and technical documents and checking samples, the product complies with the following standards currently valid in P.R.China:

GB/T 3836.1-2021, GB/T 3836.4-2021, GB/T 4208-2017

Date of Issue: 1. 23. 2024

Valid till: 1. 22. 2029



Authorize:

Coal Science (Beijing) Test Technology Co., Ltd.

(Beijing Test Center of Mine Working Equipment and Mining Oils for National Safe Production)

Note: This certificate is only valid for the products that are in accord with samples tested and verified.

Address: Building 6, No.5, Yulong Street, Caiyu Town, Beijing Economic and Technological Development Area(Daxing), Beijing Tel/fax: 010-89268018

The query url of Explosion-proof Certificate: <http://www.ccrirc.com.cn/Home>

